






- A1 THERMOMAGNETIC SWITCH N.V.
- B1 MAIN SWITCH
- B2 COOL-DOWN SWITCH
- B3 DOOR CONTACT SWITCH
- B4 SWITCH PREHEATING BOILER
- B6 CORE TEMPERATURE SENSOR SWITCH
- C1 CONTACTOR MOTOR
- C2 ROOM HEATING ELEMENTS CONTACTOR
- C4 BOILER HEATING ELEMENTS CONTACTOR
- C5 PREHEATING BOILER CONTACTOR
- E1 FUSES 0.25 A (T)
- E2 FUSES 0.125 A (T)
- E3 FUSES 0.125 A (T)
- F2 SAFETY THERMOSTAT ROOM
- F3 SAFETY THERMOSTAT BOILER
- K1 SUPPRESSION TRANSFORMERS
- K2 FILTER
- L1 ROOM LIGHT
- N MOTOR
- NA TERMINAL BLOCK
- N1 ALARM BELL
- N2 ALARM FOR CORE TEMPERATURE SENSOR
- PC PC BOARD CORE TEMPERATURE SENSOR
- PR ELECTRONIC IN-PUT (PC BOARD)
- R1 ROOM HEATING ELEMENTS
- R2 BOILER HEATING ELEMENTS
- S1 ELECTRIC VALVE HUMIDIFIER
- S2 ELECTRIC VALVE BOILER FILLING
- T1 TRANSFORMER PC-BOARD 12V
- T3 TRANSFORMER LIGHT 12V
- T6 TRANSFORMER

ACCESSORI - FITTINGS - ZUBEHÖRTEIL - ACCESSORIE - ACCESORIOS



LABORATORIO
VERIFICATO ED APPROVATO
Data 23/04/98 Firma *Paul*

05/07/90 03.0.1056
ME 40 USA 480 V